



DPD Total Chlorine Reagent Powder Pillows, 25 mL, pk/1000

Proizvod #:

1406428

Opasno po zdravlje

EUR Cijena:

Kontaktirajte nas

Dostupno



For total chlorine determination by the DPD method. USEPA accepted for reporting drinking and wastewater analysis using Hach Method 8167.

Ranges: Low Range: up to - 2.00 mg/L; Mid Range: 0.05 - 4.00 mg/L; High Range: 0.1- 10.0 mg/L. Sample Size: Low Range - 25 mL. Mid Range: 10 mL. High Range: 5 mL.

Pack of 1000 powder pillows.

Boric Acid free!

Specifikacije

Broj metode:	8167
Broj testova:	1000
Čuvanje uređaja:	10 - 25 °C
EPA compliant:	Da
Mjerni rang:	0,02 - 2,00 mg/L Cl ₂
Mjerni rang (2):	0,1 - 10,0 mg/L Cl ₂
Naziv metode:	DPD
Parametar:	Total Chlorine, Bromine, Iodine
Platforma:	Praškasti reagensi PP